



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/731,178	12/06/2000	Steven D. Goedeke	P-8896	9273

27581 7590 11/03/2003

MEDTRONIC, INC.  
710 MEDTRONIC PARKWAY NE  
MS-LC340  
MINNEAPOLIS, MN 55432-5604

EXAMINER

OPSASNICK, MICHAEL N

ART UNIT	PAPER NUMBER
----------	--------------

2655

//

DATE MAILED: 11/03/2003

Please find below and/or attached an Office communication concerning this application or proceeding.

**Office Action Summary**

Application No.

09/731,178

Applicant(s)

GOEDEKE ET AL.

Examiner

Michael N. Opsasnick

Art Unit

2655

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).
- Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 18 August 2003.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-30 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-30 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- 11) ☐ The proposed drawing correction filed on \_\_\_\_\_ is: a) ☐ approved b) ☐ disapproved by the Examiner.
- If approved, corrected drawings are required in reply to this Office action.
- 12) ☐ The oath or declaration is objected to by the Examiner.

**Priority under 35 U.S.C. §§ 119 and 120**

- 13) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- \* See the attached detailed Office action for a list of the certified copies not received.
- 14) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e) (to a provisional application).
- a) ☐ The translation of the foreign language provisional application has been received.
- 15) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. §§ 120 and/or 121.

**Attachment(s)**

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                             | 4) <input type="checkbox"/> Interview Summary (PTO-413) Paper No(s). _____  |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)         | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449) Paper No(s) _____ | 6) <input type="checkbox"/> Other: _____                                    |

Art Unit: 2655

## DETAILED ACTION

### *Drawings*

1. This application has been filed with informal drawings which are acceptable for examination purposes only. Formal drawings will be required when the application is allowed.

### *Claim Rejections - 35 USC § 103*

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. Claims 1-30 are rejected under 35 U.S.C. 103(a) as being unpatentable over Jeffery D. Snell (U.S. Patent 5,792,204, issued August 11, 1998).

As per claims 1, 4-6, 9, 19, 20, 26, and 30, Snell teaches a system interfacing with an implanted medical device (col. 3, lines 61-65), with:

- microphone input of a voice command to a speech recognizer (col. 3, lines 5-7);
- the speech recognizer matching the input voice command to the subset of commands and converting the recognized voice command into a selection code (control program instructions, col. 4, line 5),

Art Unit: 2655

- said commands along with a set of control signals being stored in memory (col. 5, line 67 thru col. 6, line 2; col. 4, lines 5-6), and
- generating a control signal therefrom to execute the commands (col. 3, lines 9-13; col. 5, lines 4-6);
- a display device (col. 4, line 62); and
- displaying received data generated by the implanted medical device in response to the execution of the command as well as implanted medical device state data (col. 4, lines 62-65 and col. 5, lines 3-13).

Snell does not explicitly teach displaying the selectable subset of commands as a function of the device. However, the examiner takes Official Notice that it is old and notoriously well known to have context-sensitive commands, and to display them for user selection by voice (or, of course, by keyboard or mouse) from a displayed menu. It would have been obvious for an artisan at the time of invention to display the available commands to avoid wasting user time by having her remember the precise available command words.

As per claims 2, 3, 23, and 29, Snell does not specifically teach a bandpass amplifier to reject ambient background signals from the microphone. However, the examiner takes Official Notice that it is old and notoriously well-known to bandpass-filter microphone audio input for speech recognizers to confine the bandwidth to the speech band. It would have been obvious for an artisan at the time of invention to do this to reduce extraneous noise, particularly broadband impulsive noise, coming from outside the speech band and causing speech recognizer errors.

As per claim 7, Snell teaches a pacing system analyzer (col. 4, lines 26-27 and 39).

As per claim 8, Snell teaches a programming unit adapted to interrogate and program the implanted medical device (col. 4, lines 62-64 and col. 5, lines 14-19).

As per claims 10-12, Snell does not explicitly teach an unidirectional microphone to be steered by the user of his medical data processing instrument. However, the examiner takes Official Notice that it is old and notoriously well-known to have a unidirectional microphone steerable by the user to input voice commands (or other speech) to a speech recognizer. It would have been obvious for an artisan at the time of invention to thus use a unidirectional microphone to conveniently spatially filter out noise coming from different directions than the user location.

As per claims 13, 24, and 27, Snell teaches adapting or configuring the speech recognizer and the processor for new commands or to a new user generating appropriate recognition data, to be stored in the memory arrangement (stored replaced command instructions and data to be used therefor, respectively, col. 5, lines 14-19 and col. 6, lines 6-9).

As per claims 14, 25, and 28, Snell does not teach validating the user to limit the various levels of commands that a user is authorized to give. However, the examiner takes Official Notice, that it is notoriously well-known in speech recognizer applications involving safety and security considerations (not to mention proprietary data or other privacy concerns) to validate the user before giving access to the corresponding command or data levels. Therefore it would have

Art Unit: 2655

been obvious for an artisan at the time of invention to include such a user validation feature in the processor, to prevent unqualified users from endangering the patient by misusing the implanted medical device (*cf.* col. 6, lines 56-58) or misappropriating her private data (*cf.* col. 5, lines 19-21).

As per claims 15-17 and 22, Snell teaches an audio signal confirming the receipt of a voice selected command or device state to inform the user or for user confirmation (col. 6, lines 61-64; col. 7, lines 4-14 and 58-62; for suggestion of similarly outputting device state *cf.* col. 5, lines 7-9).

As per claim 18, Snell teaches medical data processing from the implantable device via a communications network (col. 5, lines 41-47).

As per claim 21, Snell teaches validating the stored command speech sets to see whether the voice command is understood (col. 9, lines 6-14).

### ***Response to Arguments***

4. Applicant's arguments filed 8/18/2003 have been fully considered but they are not persuasive. As per applicant's arguments that the "examiner fails to teach a system interfacing with a medical device, and a speech recognition circuit coupled....set of commands", examiner points to Snell teaching Snell teaches a system interfacing with an implanted medical device

Art Unit: 2655

(col. 3, lines 61-65), with a microphone input of a voice command to a speech recognizer (col. 3, lines 5-7); control program instructions, col. 4, line 5. As per the arguments that no reference has been provided teachings a selectable subset of commands as a function of the device, examiner interprets this as a challenge to the Official Notice taken by the examiner, and hence points to the previously reference prior art of Brant et al (6278975) teaching the display of selectable commands in the voice recognition module (fig. 7).

### ***Conclusion***

**5. THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

**6. Any response to this action should be mailed to:**

Commissioner of Patents and Trademarks  
Washington, D.C. 20231

Art Unit: 2655

**or faxed to:**

(703) 872 9314,

(for informal or draft communications, please label "PROPOSED" or "DRAFT")

Hand-delivered responses should be brought to Crystal Park II, 2121 Crystal Drive, Arlington, VA., Sixth Floor (Receptionist).

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Opsasnick, telephone number (703)305-4089, who is available Tuesday-Thursday, 9AM-4PM.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Ms. Doris To, can be reached at (703)305-4827. The facsimile phone number for this group is (703)872-9314.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the Group 2600 receptionist whose telephone number is (703) 305-4750, the 2600 Customer Service telephone number is (703) 306-0377.

mno

AU 2655

10/25/2003

  
DORIS H. TO 10/30/03  
SUPERVISORY PATENT EXAMINER  
TECHNOLOGY CENTER 2600